

DentaPure™

DP365B Independent Water Bottle Cartridge

INSTALLATION INSTRUCTIONS



Getting Started

Refer to Instructions for Use enclosed with the cartridge and also available at HuFriedyGroup.com/DentaPure for precautionary statements, directions for use, storage, disposal and troubleshooting prior to first use of the new unit.

! Before you begin, please have the following available:

- The included DentaPure™ DP365B Cartridge installation kit containing the cartridge and luer fitting.
- Scissors or knife.

For more information on the DentaPure™ Cartridge visit:
HuFriedyGroup.com/DentaPure

For installation, or other DentaPure™ Cartridge questions, contact:

Customer Support

1-800-HuFriedy

(1-800-483-7433)

Care@Hu-Friedy.com or

DentaPureTechnical@Crosstex.com



Install your DentaPure™ Cartridge in just a few easy steps:

1 Remove

Remove the dental unit's water bottle to reveal the pick-up tube.



2 Align

Align the bottle alongside the pick-up tube to ensure your cartridge can attach to the pick-up tube with enough space to fit inside the bottle (Leave ¼" to ½" of empty space from the bottom of the bottle).



3 Cut

Once measured, cut the pick-up tube and securely insert the included luer fitting.



4 Twist

Take off the protective end caps (they will be black) and attach the cartridge to the luer fitting with a simple twist.



5 Fill

Fill the bottle with potable water and return to manifold. Check that the PSI is set to 45. Operate dental unit instruments until all air is purged from the water system.

Note: Initial dark discoloration of water is normal and is caused by the release of activated carbon. Operate water until flow appears clear.



Hu-Friedy Mfg. Co., LLC, 1666 E. Touhy Ave., Des Plaines, IL 60018 | Hu-Friedy.com

All company and product names are trademarks of Hu-Friedy Mfg. Co., LLC its affiliates or related companies, unless otherwise noted.

©2022 Hu-Friedy Mfg. Co., LLC. All rights reserved. DLIT00447 Rev A 1022