

An IMS® Product

**COLOR CODE
INSTRUMENT RINGS**

Steam sterilize for at least 4 minutes
at 270°F/132°C or 30 minutes at
250°F/121°C. Do not exceed
350°F/177°C. Recommend 30-minute
dry time after sterilization cycle.

IMS-12811 | 50 RINGS/LAVENDER

Manufactured for:
Hu-Friedy Mfg. Co., LLC
3232 N. Rockwell St.
Chicago, IL 60618
USA
Hu-Friedy.com
B3871/0518



INK COLORS



BLACK